

GB

**Radio Equipment** 

Name: M200 Fall Monitor

M200, M200-C, M210 Model(s):

**Manufacturer** 

Title:

Name: Stanley Security Systems Inc. AeroScout, LTD, transacting as STANLEY Healthcare.

STANLEY Healthcare Div a brand name and operating division of Stanley Security

Solutions Inc.

Importer:

Address: 75 Portsmouth Blvd. Stanley Black and Decker

> Suite #220 270 Bath Road Portsmouth, NH 03801 Slough, SL14DX

> > **England**

This declaration of conformity is issued under the sole responsibility of the manufacturer who declares that the above described equipment is in conformity with the following relevant Union harmonisation legislations:

The Supply of Machinery (Safety) Regulations 2016 S.I. 2016/1101 (as amended)

Electromagnetic Compatibility Regulations 2016 S1 2016/1091(as amended)

The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012, S.I. 2012/3032 (as amended)

Conformity to the essential requirements of the legislation(s) have been demonstrated by using the following **Designated standards:** 

Health and Safety (Article 3(1)(a)): EN 62368-1:2014

**EMC** (Article 3(1)(b)): EN 55022:2010/AC:2011. Class A

EN 55032:2012/AC:2013

CISPR 32:2015 EN 50130-4:2011

ANSI/UL 1069: UL 1069:2007 Ed.7 +R:23 2020

The undersigned is responsible for compilation of the technical file and makes this declaration on behalf of **Stanley Healthcare:** 

**Printed Name:** Garth Jack Date of Issue: 2/12/21

Director of Quality and Regulatory Compliance Place of Issue: Lincoln, Nebraska USA

Name of UK Representative: Kenton Madge, VP International Sales.